


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Cytokine Panel 2 (4-Plex)
Software Product Code: HA1M200518-P24P

REF Part Number: 112033HU
LOT Lot Number: HC4K200518 18793 001
Expiration Date: 2023-DEC-02

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	112052HU	HC4P211202	N/A	2023-DEC-02
Sample Diluent	101104HU	HQ6S220909	N/A	2024-SEP-09
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M200518	650 µL	2024-DEC-15
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220407	N/A	2024-APR-07
Streptavidin HRP	101173GR	QNSN221024	N/A	2024-APR-24
Substrate A	101159GR	QNSA221020	N/A	2024-OCT-20
Substrate B+	101120GR	QNSB+221020	N/A	2024-OCT-20
Wash Buffer	101158GR	QNSW221123	N/A	2024-NOV-23

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1α	pg/mL	3900	3900	5.35	2.67
2	IL-4	pg/mL	1600	1600	2.19	1.10
3	IL-6	pg/mL	3100	3100	4.25	2.13
4	TNFα	pg/mL	5400	5400	7.41	3.70
5	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Stacey Braslow

Date: 15 DEC 22

QA Release: Sylvia Adams

Date: 15 DEC 22

for Research Use Only