

CERTIFICATE OF ANALYSIS

 Q-Plex™ Human Cytokine Panel 1 (6-Plex)
 Software Product Code: HA1M200518-P16P


 REF Part Number: 115233HU

 LOT Lot Number: HCP1K200518 19227 001

 Expiration Date: 2024-APR-07

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	115252HU	HCP1P220412	N/A	2024-APR-12
Sample Diluent	101104HU	HQ6S220909	N/A	2024-SEP-09
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M200518	650 µL	2025-APR-14
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6220407	N/A	2024-APR-07
Streptavidin HRP	101173GR	QNSN221228	N/A	2024-JUN-28
Substrate A	101159GR	QNSA221221	N/A	2024-DEC-21
Substrate B+	101120GR	QNSB+221221	N/A	2024-DEC-21
Wash Buffer	101158GR	QNSW230314	N/A	2025-MAR-14

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1β	pg/mL	11800	11800	16.19	8.09
2	IL-6	pg/mL	3100	3100	4.25	2.13
3	TNFα	pg/mL	5400	5400	7.41	3.70
4	IFNγ	pg/mL	8000	8000	10.97	5.49
5	IL-1α	pg/mL	3900	3900	5.35	2.67
6	IL-10	pg/mL	6800	6800	9.33	4.66
7	REF	N/A	N/A	N/A	N/A	N/A


Q-Plex™ ARRAYS

QC Release:

Date:

17 APR 23

QA Release:

Date:

17 APR 23

for Research Use Only