

101896GR - 6.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (14-Plex)
Software Product Code: HPNM007SPWHO14-CP

REF Part Number: 481849HU
LOT Lot Number: HPNK220926 19460 001
Expiration Date: 2023-SEP-26

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481852HU	HPNP230501	N/A	2024-MAY-01
Sample	481157HU	HPNS220926	N/A	2023-SEP-26
Calibrator	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM220926	500 µL	2024-JUN-13
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND221027	N/A	2023-OCT-27
Streptavidin	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA230302	N/A	2025-MAR-02
Substrate B+	101120GR	QNSB+230302	N/A	2025-MAR-02
Wash Buffer	101158GR	QNSW230314	N/A	2025-MAR-14

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	1	µg/mL	0.850	0.148	0.00027	0.00013
2	3	µg/mL	0.145	0.102	0.00019	0.000093
3	4	µg/mL	0.333	0.058	0.00010	0.000052
4	5	µg/mL	0.751	0.132	0.00024	0.00012
5	6B	µg/mL	0.905	0.157	0.00029	0.00015
6	7F	µg/mL	0.830	0.146	0.0011	0.00053
7	8	µg/mL	1.424	0.249	0.0018	0.00090
8	9N	µg/mL	0.703	0.125	0.00089	0.00045
9	9V	µg/mL	0.644	0.112	0.00079	0.00039
10	12F	µg/mL	0.221	0.038	0.00028	0.00014
11	14	µg/mL	3.799	0.165	0.00121	0.00060
12	18C	µg/mL	0.730	0.129	0.00093	0.00047
13	19F	µg/mL	1.461	0.256	0.0019	0.00094
14	23F	µg/mL	0.595	0.102	0.00076	0.00038
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]
QA Release: [Signature]
for Research Use Only

Date: 14 AUG 23
Date: 14 AUG 23