

101896GR - 6.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Angiogenesis (9-Plex)
Software Product Code: HA1M220712-AG

REF Part Number: 150233HU
LOT Lot Number: HAGK220712 19510 001
Expiration Date: 2024-AUG-16

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	150252HU	HAGP230424	N/A	2025-APR-24
Sample Diluent	101104HU	HQ6S220909	N/A	2024-SEP-09
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2025-JUN-26
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	150256HU 01	HAGDQ6221230	N/A	2024-DEC-30
Streptavidin HRP	101173GR	QNSN230216	N/A	2024-AUG-16
Substrate A	101159GR	QNSA230302	N/A	2025-MAR-02
Substrate B+	101120GR	QNSB+230302	N/A	2025-MAR-02
Wash Buffer	101158GR	QNSW230314	N/A	2025-MAR-14

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	ANG-2	pg/mL	8460	8460	11.60	5.80
2	FGF	pg/mL	4950	4950	6.79	3.40
3	HGF	pg/mL	15160	15160	20.80	10.40
4	IL-8	pg/mL	2080	2080	2.85	1.43
5	PDGF	pg/mL	4100	4100	5.62	2.81
6	TIMP-1	pg/mL	12840	12840	52.84	26.42
7	TIMP-2	pg/mL	19135	19135	26.25	13.12
8	TNFα	pg/mL	6020	6020	8.26	4.13
9	VEGF	pg/mL	1890	1890	2.59	1.30
10	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]
QA Release: [Signature]
for Research Use Only

Date: 26 JUN 23
Date: 26 JUN 23