

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (14-Plex)
Software Product Code: HPNM007SPARUP14-CP



REF Part Number: 481849HU
LOT Lot Number: HPNK220926 19561 001
Expiration Date: 2024-JUN-14

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481852HU	HPNP230614	N/A	2024-JUN-14
Sample Diluent	481157HU	HPNS230621	N/A	2024-JUN-21
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM220926	500 µL	2024-JUL-06
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND230622	N/A	2024-JUN-22
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA230302	N/A	2025-MAR-02
Substrate B+	101120GR	QNSB+230302	N/A	2025-MAR-02
Wash Buffer	101158GR	QNSW230314	N/A	2025-MAR-14

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	1	1.49	µg/mL	0.259	0.000468	0.000234
2	3	0.42	µg/mL	0.294	0.000533	0.0002670
3	4	0.88	µg/mL	0.154	0.000273	0.0001370
4	5	2.12	µg/mL	0.371	0.000676	0.000338
5	6B	0.77	µg/mL	0.133	0.000247	0.000124
6	7F	2.22	µg/mL	0.392	0.00286	0.001430
7	8	0.76	µg/mL	0.133	0.000962	0.000481
8	9N	0.87	µg/mL	0.154	0.00111	0.000553
9	9V	1.17	µg/mL	0.203	0.00143	0.000715
10	12F	0.41	µg/mL	0.070	0.000520	0.000260
11	14	2.74	µg/mL	0.119	0.000871	0.000436
12	18C	1.22	µg/mL	0.217	0.00156	0.000780
13	19F	5.28	µg/mL	0.924	0.00676	0.003380
14	23F	0.49	µg/mL	0.084	0.000624	0.000312
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]

Date: 11 JUL 23

QA Release: [Signature]

Date: 11 JUL 23

for Research Use Only