

101896GR - 7.0

# QUANSYS BIOSCIENCES CERTIFICATE OF ANALYSIS

Quansys Biosciences  
365 N 600 W Logan, UT 84321  
(435) 752-0531

Q-Plex™ Porcine Cytokine HS (4-Plex)  
Software Product Code: PCSM170306CY




**REF** Part Number: 119149PC  
**LOT** Lot Number: PCSK170306 19675 002  
Expiration Date: 2024-DEC-28

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	119152PC	PCSP230609	N/A	2025-JUN-09
Sample Diluent	119157PC	PCSS230613	N/A	2025-JUN-13
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	119155PC	PCSM170306	500 µL	2025-AUG-08
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	119156PC	PCSD230615	N/A	2025-JUN-15
Streptavidin HRP	101173GR	QNSN230628	N/A	2024-DEC-28
Substrate A	101159GR	QNSA230602	N/A	2025-JUN-02
Substrate B+	101120GR	QNSB+230602	N/A	2025-JUN-02
Wash Buffer	101158GR	QNSW230718	N/A	2025-JUL-18

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-1β	1600	pg/mL	N/A	N/A	2.10
2	IL-6	2000	pg/mL	N/A	N/A	2.50
3	IL-8	2300	pg/mL	N/A	N/A	2.90
4	TNFα	1350	pg/mL	N/A	N/A	1.90
5	REF	N/A	N/A	N/A	N/A	N/A

Q-Plex™ ARRAYS



QC Release: [Signature]  
QA Release: [Signature]

Date: 09 AUG 23

Date: 09 AUG 23

for Research Use Only