

101896GR - 7.0

QUANSYS BIOSCIENCES

CERTIFICATE OF ANALYSIS

Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


Q-Plex™ Human Cytokine Panel 1 (6-Plex)
Software Product Code: HA1M220712-P16P



REF Part Number: 115233HU
LOT Lot Number: HCP1K220712 19796 001
Expiration Date: 2024-JUN-28

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	115252HU	HCP1P230605	N/A	2025-JUN-05
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HAIM220712	600 µL	2025-SEP-01
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6230612	N/A	2025-JUN-12
Streptavidin HRP	101173GR	QNSN221228	N/A	2024-JUN-28
Substrate A	101159GR	QNSA230602	N/A	2025-JUN-02
Substrate B+	101120GR	QNSB+230602	N/A	2025-JUN-02
Wash Buffer	101158GR	QNSW230801	N/A	2025-AUG-01

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-1β	11800	pg/mL	11800	16.19	8.09
2	IL-6	2850	pg/mL	2850	3.91	1.95
3	TNFα	6020	pg/mL	6020	8.26	4.13
4	IFNγ	9775	pg/mL	9775	13.41	6.70
5	IL-1α	3940	pg/mL	3940	5.40	2.70
6	IL-10	8400	pg/mL	8400	11.52	5.76
7	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: *J. Wa*
QA Release: *[Signature]*
for Research Use Only

Date: 08 SEP 23
Date: 08 SEP 23