

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Custom 4-Plex
Software Product Code: HA1M220712-ID4



REF Part Number: 107749GR
LOT Lot Number: HCUK220712 19900 001
Expiration Date: 2024-JUL-13

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP230919	N/A	2025-SEP-19
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2025-SEP-25
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6230612	N/A	2025-JUN-12
Streptavidin HRP	101173GR	QNSN230628	N/A	2024-DEC-28
Substrate A	101159GR	QNSA220713	N/A	2024-JUL-13
Substrate B+	101120GR	QNSB+220713	N/A	2024-JUL-13
Wash Buffer	101158GR	QNSW230801	N/A	2025-AUG-01

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-1β	11800	pg/mL	11800	16.19	8.09
2	IL-6	2850	pg/mL	2850	3.91	1.95
3	IL-8	2080	pg/mL	2080	2.85	1.43
4	TNFα	6020	pg/mL	6020	8.26	4.13
5	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: *[Signature]*
 QA Release: *[Signature]*
 for Research Use Only

Date: 20SEP23

Date: 20SEP23