

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex) Extra Diluent
Software Product Code: HPNM230726-WHO9 



REF Part Number: 481449HU
LOT Lot Number: HPNK230726 20018 001
Expiration Date: 2024-JUL-26

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP230726	N/A	2024-JUL-26
Sample Diluent	481157HU-40	HPNS231011	N/A	2024-OCT-11
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM230726	500 µL	2024-OCT-25
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND230730	N/A	2024-JUL-30
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA2310720	N/A	2025-JUL-20
Substrate B+	101120GR	QNSB+230720	N/A	2025-JUL-20
Wash Buffer	101158GR	QNSW230718	N/A	2025-JUL-18

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	2	2.46	µg/mL	0.43	0.0031	0.0016
2	10A	1.30	µg/mL	0.057	0.0017	0.00085
3	11A	0.51	µg/mL	0.091	0.00060	0.00030
4	15B	1.69	µg/mL	0.29	0.0021	0.0011
5	17F	0.85	µg/mL	0.14	0.0043	0.0022
6	19A	1.39	µg/mL	0.25	0.0073	0.0037
7	20	1.05	µg/mL	0.18	0.0014	0.00070
8	22F	0.95	µg/mL	0.17	0.0048	0.0024
9	33F	1.07	µg/mL	0.047	0.0014	0.00070
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: 
QA Release: 

Date: 26 OCT 23

Date: 26 OCT 23

for Research Use Only