

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Chemokine (9-Plex)
Software Product Code: HA1M220712-CK



REF Part Number: 120233HU
LOT Lot Number: HCKK220712 20008 001
Expiration Date: 2024-AUG-22

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	140252HU	HCKP220822	N/A	2024-AUG-22
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2025-NOV-01
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	120256HU-01	HCKDQ6231031	N/A	2025-OCT-31
Streptavidin HRP	101173GR	QNSN230906	N/A	2025-MAR-06
Substrate A	101159GR	QNSA231006	N/A	2025-MAR-13
Substrate B+	101120GR	QNSB+231006	N/A	2025-MAR-13
Wash Buffer	101158GR	QNSW230801	N/A	2025-AUG-01

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	Eotaxin	1740	pg/mL	1740	2.39	1.19
2	GROα	1910	pg/mL	1910	2.62	1.31
3	I-309	2350	pg/mL	2350	3.22	1.61
4	IL-8	2080	pg/mL	2080	2.85	1.43
5	IP-10	3240	pg/mL	1080	4.44	2.22
6	MCP-1	1720	pg/mL	573	7.08	3.54
7	MCP-2	585	pg/mL	585	7.22	3.61
8	RANTES	6310	pg/mL	6310	8.66	4.33
9	TARC	1415	pg/mL	1415	1.94	0.97
10	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]
QA Release: [Signature]

Date: 03 NOV 23
Date: 03 NOV 23

for Research Use Only