

101896GR - 7.0



Quansys Biosciences  
365 N 600 W Logan, UT 84321  
(435) 752-0531

**CERTIFICATE OF ANALYSIS**


Q-Plex™ Mouse Cytokine Panel 2 HS (4-Plex)  
Software Product Code: MHSM230927-P24P



**REF** Part Number: 111149MS  
**LOT** Lot Number: MCP4K230927 20247 001  
Expiration Date: 2024-JUN-30

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	111152MS	MCP4P231012	N/A	2024-OCT-12
Sample Diluent	112457MS	MHSS230512	N/A	2024-JUN-30
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	112455MS	MHSM230927	550 µL	2024-DEC-19
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	112456MS	MHSD231013	6 mL	2024-OCT-13
Streptavidin HRP	101173GR	QNSN231025	N/A	2025-APR-25
Substrate A	101159GR	QNSA231205	N/A	2025-MAR-13
Substrate B+	101120GR	QNSB+231205	N/A	2025-MAR-13
Wash Buffer	101158GR	QNSW231024	N/A	2025-OCT-24

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IFN $\gamma$	800	pg/mL	800	1.10	0.55
2	IL-1 $\beta$	2390	pg/mL	2390	3.28	1.64
3	IL-6	1110	pg/mL	1110	1.52	0.76
4	TNF $\alpha$	390	pg/mL	390	0.53	0.27
5	REF	N/A	N/A	N/A	N/A	N/A



**Q-Plex™ ARRAYS**

QC Release:

*[Signature]*

Date:

19 DEC 23

QA Release:

*[Signature]*

Date:

19 DEC 23

for Research Use Only