

101896GR - 7.0

**QUANSYS**  
BIOSCIENCES  
**CERTIFICATE OF ANALYSIS**


Quansys Biosciences  
365 N 600 W Logan, UT 84321  
(435) 752-0531

Q-Plex™ Human Cytokine Screen (16-Plex)  
Software Product Code: HA1M220712-CY

**REF** Part Number: 110933HU  
**LOT** Lot Number: HCSK220712 20597 001  
Expiration Date: 2025-MAY-25

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	110952HU	HCSP230626	N/A	2025-JUN-26
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2026-MAR-22
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6230612	N/A	2025-JUN-12
Streptavidin HRP	101173GR	QNSN240103	N/A	2025-JUL-03
Substrate A	101159GR	QNSA231027	N/A	2025-AUG-23
Substrate B+	101120GR	QNSB+231027	N/A	2025-AUG-23
Wash Buffer	101158GR	QNSW240208	N/A	2026-FEB-08

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-1α	3940	pg/mL	3940	5.40	2.70
2	IL-1β	11800	pg/mL	11800	16.19	8.09
3	IL-2	5150	pg/mL	5150	7.06	3.53
4	IL-4	1670	pg/mL	1670	2.29	1.15
5	IL-5	2950	pg/mL	2950	4.05	2.02
6	IL-6	2850	pg/mL	2850	3.91	1.95
7	IL-8	2080	pg/mL	2080	2.85	1.43
8	IL-10	8400	pg/mL	8400	11.52	5.76
9	IL-12p70	2890	pg/mL	2890	3.96	1.98
10	IL-13	2865	pg/mL	2865	3.93	1.97
11	IL-15	9800	pg/mL	9800	13.44	6.72
12	IL-17	19500	pg/mL	19500	26.75	13.37
13	IL-23	45950	pg/mL	45950	63.03	31.52
14	IFNγ	9775	pg/mL	9775	13.41	6.70
15	TNFα	6020	pg/mL	6020	8.26	4.13
16	TNFβ	4200	pg/mL	4200	5.76	2.88
17	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]  
QA Release: [Signature]

Date: 26 MAR 24  
Date: 26 MAR 24

for Research Use Only