


101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Custom 4-Plex
Software Product Code: HA1M220712-EV4 



REF Part Number: 107749GR-04
LOT Lot Number: HCUK220712 20708 001
EXP Expiration Date: 2024-OCT-13

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP240413	N/A	2026-APR-13
Sample Diluent	101104HU	HQ6S240125	N/A	2026-JAN-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2026-APR-23
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6230612	N/A	2025-JUN-12
Streptavidin HRP	101173GR	QNSN240103	N/A	2025-JUL-03
Substrate A	101159GR	QNSA231205	N/A	2024-OCT-13
Substrate B+	101120GR	QNSB+231205	N/A	2024-OCT-13
Wash Buffer	101158GR	QNSW240208	N/A	2026-FEB-08

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-8	2080	pg/mL	2080	2.85	1.43
2	IL-10	8400	pg/mL	8400	11.52	5.76
3	IL-17a	19500	pg/mL	19500	26.75	13.37
4	TNFa	6020	pg/mL	6020	8.26	4.13
5	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: 
QA Release: 
for Research Use Only

Date: 24 APR 24
Date: 24 APR 24