

101896GR - 7.0

QUANSYS
BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531

Q-Plex™ Human Pneumococcal IgG (14-Plex)
Software Product Code: HPNM231211-W14P

REF Part Number: 481849HU
LOT Lot Number: HPNK231211 21071 001
Expiration Date: 2024-NOV-27

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481852HU	HPNP231127	N/A	2024-NOV-27
Sample Diluent	481157HU	HPNS240401	N/A	2025-APR-01
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM231211	500 µL	2025-JUL-23
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND231128	N/A	2024-NOV-28
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA240405	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240405	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW240412	N/A	2026-APR-12

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	1	0.76	µg/mL	0.19	0.00075	0.00037
2	3	0.12	µg/mL	0.12	0.00048	0.00024
3	4	0.29	µg/mL	0.073	0.00028	0.00014
4	5	0.74	µg/mL	0.18	0.00072	0.00036
5	6B	0.85	µg/mL	0.21	0.00083	0.00041
6	7F	0.77	µg/mL	0.19	0.0030	0.0015
7	8	1.30	µg/mL	0.33	0.0051	0.0025
8	9N	0.68	µg/mL	0.17	0.0026	0.0013
9	9V	0.58	µg/mL	0.14	0.0023	0.0011
10	12F	0.20	µg/mL	0.050	0.00077	0.00039
11	14	3.43	µg/mL	0.21	0.0033	0.0017
12	18C	0.66	µg/mL	0.17	0.0026	0.0013
13	19F	1.24	µg/mL	0.31	0.0048	0.0024
14	23F	0.53	µg/mL	0.13	0.0021	0.0010
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature] Date: 25 JUL 24
 QA Release: [Signature] Date: 25 JUL 24
 for Research Use Only