

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex) Kit 30 mL Diluent
Software Product Code: HPNM240514-A9P

REF Part Number: 481249HU
Lot Number: HPNK240514 20935 001
Expiration Date: 2025-JUN-11

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP240618	N/A	2025-JUN-18
Sample Diluent	481157HU-30	HPNS240611	N/A	2025-JUN-11
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM240514	500 µL	2025-JUN-27
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND240621	N/A	2025-JUN-21
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA240531	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240531	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW240514	N/A	2026-MAY-14

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	2	2.19	µg/mL	0.43	0.0031	0.0016
2	10A	1.16	µg/mL	0.057	0.0017	0.00085
3	11A	0.44	µg/mL	0.091	0.00060	0.00030
4	15B	1.51	µg/mL	0.29	0.0021	0.0011
5	17F	0.74	µg/mL	0.14	0.0043	0.0022
6	19A	1.36	µg/mL	0.25	0.0073	0.0037
7	20	0.91	µg/mL	0.18	0.0014	0.00070
8	22F	0.85	µg/mL	0.17	0.0048	0.0024
9	33F	0.96	µg/mL	0.047	0.0014	0.00070
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature] Date: 27 JUN 24
 QA Release: [Signature] Date: 27 JUN 24

for Research Use Only