


101896GR - 8.0

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS

Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Pneumococcal IgG (9-Plex) Aliquotted Bulk 30 mL Diluent

REF	Part Number: 481249HU-25	Software Product Code: HPNM240514-A9P 
LOT	Lot Number: HPNK240514 21560 001	
EXP	Expiration Date: 2025-JUN-11	



Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP240618	N/A	2025-JUN-18
Sample Diluent	481157HU-30	HPNS240611	N/A	2025-JUN-11
Calibrator	481155HU	HPNM240514	500 µL	2025-NOV-14
Detection	481156HU	HPND240621	N/A	2025-JUN-21
Substrate A	101159GR	QNSA240531	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240531	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW240925	N/A	2026-SEP-25



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	2	2.19	µg/mL	0.0016	0.0031	0.43
2	10A	1.16	µg/mL	0.00085	0.0017	0.057
3	11A	0.44	µg/mL	0.00030	0.00060	0.091
4	15B	1.51	µg/mL	0.0011	0.0021	0.29
5	17F	0.74	µg/mL	0.0022	0.0043	0.14
6	19A	1.36	µg/mL	0.0037	0.0073	0.25
7	20	0.91	µg/mL	0.00070	0.0014	0.18
8	22F	0.85	µg/mL	0.0024	0.0048	0.17
9	33F	0.96	µg/mL	0.00070	0.0014	0.047
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	REF	N/A	N/A	N/A	N/A	N/A

QC Release:  Date: 14 NOV 24
 QA Release:  Date: 14 NOV 24
 for Research Use Only